



H₂S GUARDIAN GAS DETECTOR

Battery Replacement Procedure:

⚠️WARNING! DO NOT SERVICE DEVICE IN A HAZARDOUS ENVIRONMENT⚠️

To replace the battery, follow these steps:

1. Ensure you are not in a hazardous environment.
2. Unscrew 4 corner screws on back of device with a Phillips screwdriver.
3. Carefully separate top case from back, ensure the housing gasket does not come off of corner posts.
4. Remove existing Panasonic CR123A battery.
5. Insert new Panasonic CR123A battery (ensure correct battery polarity).
6. Ensure housing gasket is in place and screw housing halves together. Torque to 30 in-oz.
7. Bump test device to ensure sensor gasket did not get misaligned. If bump test fails, re-calibrate or re-open device to ensure gaskets are aligned properly.

Please contact Graphic Controls with any questions about this procedure or to learn more about our factory calibration service program at 1-800-669-1535 or www.graphiccontrols.com.